



Risperidone Injection Incident Report

From time to time pharmacy receives reports of problems occurring during the reconstitution of risperidone long acting injection, RLAI, (Risperdal Consta[®]). The manufacturers will refund or replace faulty injections when we return them.

Please report any difficulties that you experience or encounter when using RLAI.

Team/Unit Name	
Date	
Patient's registration (ePEX) No.	
Strength of RLAI	25mg / 37.5mg / 50mg
Batch Number (Enter the whole of all the numbers on the outer box)	
Expiry date (Taken from the outer box)	
Nature of fault	
During reconstitution all the diluent leaked out between the vial access device and the vial	Yes / No
Part way through reconstitution resistance was observed; some diluent entered the vial some leaked out	Yes / No
The white collar came off the syringe during the reconstitution process	Yes / No
The plunger handle came away from the rubber plunger during the reconstitution process	Yes / No
The white collar came off the syringe during the administration process	Yes / No
The plunger handle came away from the rubber plunger during the administration process	Yes / No
The needle unscrewed after reconstitution and before administration.	Yes / No
Other, please specify	
Has injection been returned with this form?	Yes / No
Has injection been used, albeit with difficulty?	Yes / No
Name and Signature of nurse	